

New Delhi, the 7th September, 2001

S.O. 2473.—Whereas the Central Government, after considering the report submitted to it by the prescribed authority, is satisfied that the model described in the said report (the figure given below) is in conformity with the provisions of the Standards of Weights and Measures Act, 1976 (60 of 1976) and the Standards of Weights and Measures (Approval of Models) Rules, 1987 and the said model is likely to maintain its accuracy over periods of sustained use and to render accurate service under varied conditions;

Now, therefore, in exercise of the powers conferred by sub-section (7) of Section 36 of the said Act, the Central Government hereby publishes the certificate of approval of the model of the self-indicating, non-automatic, (Table top piece counting type) weighing instrument with digital indication of "CT" series of High accuracy (Accuracy class II) and with brand name "EAGLE" (hereinafter referred to as the model), manufactured by M/s. E.G. Kantawalla Private Limited, 28/1, Old Nagar, Mundhwa Road, Damadar Nagar, Off : Nagar Road, Kharadi, Pune-411 014 and which is assigned the approval mark IND/09/2001/76;

The said model (figure given) is a weighing instrument with a maximum capacity of 12 kg and minimum capacity of 20 g. The verification scale interval (e) is 1g. It has a tare device with a 100 per cent subtractive retained tare effect. The Light Emitting Diode display indicates the weighing result. The instrument operates on 230 Volts and 50-Hertz alternate current power supply;



Further, in exercise of the powers conferred by sub-section (12) of Section 36 of the said Act, the Central Government hereby declares that this certificate of approval of the model shall also cover the weighing instrument of similar make and performance of same series with maximum capacity upto 50 kg and with number of verification scale interval(n) in the range of 100 to 1,00,000 for 'e' value of 1mg to 50 mg and with number of verification scale interval(n) in the range of 5,000 to 1,00,000 for 'e' value of 100 mg or more and with 'e' value 1×10^k , 2×10^k , 5×10^k , k being the positive or negative whole number or equal to zero, manufactured by the same manufacturer in accordance with the same principle, design and with the same materials with which, the approved model have been manufactured.

[F.No. WM-21(12)/2001]

P.A. KRISHNAMOORTHY, Director, Legal Metrology