

New Delhi, the 3rd December, 2004

S.O. 3279.—Whereas the Central Government, after considering the report submitted to it by the prescribed authority, is satisfied that the model described in the said report (see the figure given below), is in conformity with the provisions of the Standards of Weights and Measures Act, 1976 (60 of 1976) and the Standards of Weights and Measures (Approval of Models) Rules, 1987 and the said Model is likely to maintain its accuracy over periods of sustained use and to render accurate service under varied conditions;

Now, therefore, in exercise of the powers conferred by Sub-sections (7) and (8) of Section 36 of the said Act, the Central Government hereby issues and publishes the certificate of approval of the model of self-indicating, non-automatic, (Table top type) weighing instrument with digital indication of "JEWEL" series of high accuracy (Accuracy class-II) and with brand name "EAGLE" (herein after referred to as the said model), manufactured by M/s Shanghai Yousheng Weighing Apparatus Co. Limited, Room No. 201, Bldg. No. 10 No. 45, Zhuan Xin Road, Minhang, Shanghai, China and imported in India by M/s E. G. Kantawalla Pvt. Limited, Survey No. 28/1, Damodar Nagar, Old Nagar Mundhwa Road Kharadi, Pune—411014 and which is assigned the approval mark IND/09/2004/232,

The said Model (see the figure given below) is strain gauge type load cell based non-automatic weighing instrument (Table top type) with a maximum capacity of 300g and minimum capacity of 500mg. The verification scale interval (e) is 10mg. It has a tare device with a 100 per cent subtractive retained tare effect. The Light Emitting Diode (LED) indicates the weighing result. The instrument operates on 230 Volts and 50-hertz alternate current power supply;



In addition to sealing the stamping plate sealing shall also be done to prevent the opening of the machine for fraudulent practices.

Further, in exercise of the powers conferred by Sub-section (12) of Section 36 of the said Act, the Central Government hereby declares that this certificate of approval of the said model shall also cover the weighing instruments of similar make, and performance of same series with maximum capacity upto 50kg and with number of verification scale interval (n) in the range of 100 to 5,000 for 'e' value of 1mg to 50mg and with number of verification scale interval (n) in the range of 5,000 to 50,000 for 'e' value of 100mg or more and with 'e' value of 1×10^k , 2×10^k or 5×10^k , k being the positive or negative whole number or equal to zero manufactured by the same manufacturer in accordance with the same principle, design and with the same materials with which, the said approved model has been manufactured.

[F. No. WM-21(111)/2003]
P. A. KRISHNAMOORTHY, Director of Legal Metrology